

Bromegrass

919 Moisture Meter Chart

Instructions

- Sample size 50 grams
- Calibrate meter to 33
- Chart does not compensate for temperature.
 - Seed temperature must be Temp 19°C - 25°C (66°F - 78°F) to assess moisture accurately.
- Dry is 12%

Dial Reading	% Moisture
45	6.00
46	6.39
47	6.78
48	7.17
49	7.56
50	7.96
51	8.35
52	8.74
53	9.13
54	9.52
55	9.91
56	10.30
57	10.69
58	11.08
59	11.48
60	11.87
61	12.26
62	12.65
63	13.04
64	13.43
65	13.82
66	14.21
67	14.60
68	18.93
69	19.27
70	19.60

 Dry