

Tall Fescue & Meadow Fescue 919 Moisture Meter Chart

Instructions

- Sample size 125 grams
- Calibrate meter to 87
- Chart does not compensate for temperature.
 - Seed temperature must be Temp 19°C - 25°C (66°F - 78°F) to assess moisture accurately.
- Dry is 12%

Dial Reading	% Moisture	Dial Reading	% Moisture	Dial Reading	% Moisture	Dial Reading	% Moisture	Dial Reading	% Moisture
1	6.50	11	9.70	21	12.90	31	16.20	41	19.40
2	6.80	12	10.00	22	13.30	32	16.50	42	19.70
3	7.10	13	10.40	23	13.60	33	16.80	43	20.10
4	7.40	14	10.70	24	13.90	34	17.20	44	20.40
5	7.80	15	11.00	25	14.20	35	17.50	45	20.70
6	8.10	16	11.30	26	14.60	36	17.80	46	21.00
7	8.40	17	11.60	27	14.90	37	18.10	47	21.40
8	8.70	18	12.00	28	15.20	38	18.50	48	21.70
9	9.10	19	12.30	29	15.50	39	18.80	49	22.00
10	9.40	20	12.60	30	15.90	40	19.10	50	22.30

 Dry