

Timothy 919 Moisture Meter Chart

Instructions

- Sample size 250 grams
- Calibrate meter to 53
- Chart does not compensate for temperature.
 - Seed temperature must be Temp 19°C - 25°C (66°F - 78°F) to assess moisture accurately.
- Dry is 10%

Dial Reading	% Moisture	Dial Reading	% Moisture	Dial Reading	% Moisture
10	8.40	32	11.74	54	15.09
11	8.55	33	11.90	55	15.24
12	8.70	34	12.05	56	15.39
13	8.86	35	12.20	57	15.54
14	9.01	36	12.35	58	15.70
15	9.16	37	12.50	59	15.85
16	9.31	38	12.66	60	16.00
17	9.46	39	12.81		
18	9.62	40	12.96		
19	9.77	41	13.11		
20	10.92	42	13.26		
21	10.07	43	13.42		
22	10.22	44	13.57		
23	10.38	45	13.72		
24	10.53	46	13.87		
25	10.68	47	14.02		
26	10.83	48	14.18		
27	10.98	49	14.33		
28	11.14	50	14.48		
29	11.29	51	14.63		
30	11.44	52	14.78		
31	11.59	53	14.		

 Dry